




Volume 12, Number 6, June 2006

## Simple Respiratory Mask

Virginia M. Dato,\* David Hostler,\* and Michael E. Hahn\*

\*University of Pittsburgh, Pittsburgh, Pennsylvania, USA



[Back to article](#)

**Figure.** Prototype mask. A) Side view, B) Face side. This mask consisted of 1 outer layer ( $\approx 37 \text{ cm} \times 72 \text{ cm}$ ) rolled and cut as in panel B with 8 inner layers ( $\leq 18 \text{ cm}^2$ ) placed inside (against the face). The nose slit was first placed over the bridge of the nose, and the roll was tied below the back of the neck. The area around the nose was adjusted to eliminate any leakage. If the seal was not tight, it was adjusted by adding extra material under the roll between the cheek and nose or by pushing the rolled fabric above or below the cheekbone. Tie b was tied over the head. A cloth extension was added if tie b was too short. Finally, tie c was tied behind the head. The mask was then fit tested.

This page last reviewed May 12, 2006